

Automatic Blood Pressure Monitor



Instruction manual
Gebrauchsanweisung
Gebruiksaanwijzing

page 1 Mode d'emploi
Seite 23 Manuale di istruzioni
pagina 45 Manual de instrucciones

page 67
pagina 89
pagina 111

Automatic Blood Pressure Monitor



Instruction manual
Gebrauchsanweisung
Gebruiksaanwijzing

page 1 Mode d'emploi
Seite 23 Manuale di istruzioni
pagina 45 Manual de instrucciones

page 67
pagina 89
pagina 111

Table of Contents

Before measuring your Blood Pressure	3
Notes on safety	4
Names and functions of the parts	5
How to insert the batteries	6
How to set the Date and Time	7
How to measure Blood Pressure correctly	7
How to obtain a correct posture	9
How to measure Blood Pressure	10
How to use the Memory function	11
How to use the Graph function	12
How to change the settings	13
Auto memory	14
Position sensor	15
Left or right wrist measurement	15
Alarm	15
Time	16
Delete data	16
Reset settings	16
What is Blood Pressure?	17
Blood pressure at the upper arm and wrist	18
Maintenance and storage	19
Error indication	20
Troubleshooting	21
Spare parts and accessories	22
Specifications	22

Dear Customer,


Thank you for purchasing this OMRON product. OMRON products are known for their reliability and advanced technology. By buying this OMRON R7 wrist blood pressure monitor with position sensor you have chosen for a high quality and innovative instrument. We hope that you will be fully satisfied with your new blood pressure monitor. Please take your time to read this instruction manual carefully before using it for the first time. If you still have questions please contact your OMRON dealer or the distributor as mentioned on the packaging. They will be pleased to help you.

We wish you a healthy future!







Before measuring your Blood Pressure





Notes on Safety Please read thoroughly before using this monitor.

- The warning signs and the sample icons shown here are listed for you to use the product safely and correctly as well as to prevent the risk of damage to you and others.
- The icons and meanings are shown as follows.

Warning sign	Contents
 Caution	Indicates matters in which bodily harm or material damage* may arise as a result of incorrect handling.

* Material damage refers to a wide range of damage involving your house, household goods, domestic animals, and pets.

Examples of signs	
	The  icon indicates prohibitions (what you cannot do). Matters involving actual prohibitions are indicated by text or pictures in or near  . The left icon refers to "prohibition to disassemble".
	The  icon indicates something compulsory (always follow). Matters involving actual compulsory actions are indicated by text or pictures in or near  . The left icon refers to "general compulsion".

 Caution	
<p>Self-diagnosis of measured results and treatment are dangerous. Please follow the instructions of your doctor.</p> <ul style="list-style-type: none"> • Self-diagnosis may worsen the disease. • When the arterial sclerosis proceeds due to diabetics, hyperlipemia, or hypertension, it may cause peripheral circulatory disturbance. For these cases, blood pressure may differ greatly between measurement at the wrist and the brachial. 	
<p>Do not use this monitor for the infant or the person who cannot express one's intention.</p> <ul style="list-style-type: none"> • They may suffer accident or trouble. Do not use this monitor for purposes other than measuring blood pressure. • You may suffer accident or trouble. Do not use portable phones near this monitor. • The monitor may operate erroneously. 	
<p>Do not disassemble or modify the main unit or the wrist cuff of the blood pressure monitor.</p> <ul style="list-style-type: none"> • Correct measurement will be impossible. 	

Notes on safety

Important instructions for obtaining meaningful readings

- Avoid eating and drinking (alcohol), smoking, sport and taking medicines before measuring your blood pressure as this could affect your blood pressure level.
- Before you start sit comfortably on a chair and relax. It is recommended to check your blood pressure at least twice a day, in the morning after getting up and in the evening after work, or as advised by your doctor.
- Do not move and do not speak while the measurement is being taken. Avoid measuring your blood pressure while you are in a moving vehicle.
- Allow an interval of at least three minutes between readings. Otherwise congestion of blood might occur and distort the readings. Always remove the cuff between readings.
- Bare your wrist and take off your watch and jewellery before applying the cuff. Do not constrict the flow of your blood if you push up your sleeve.

Cautions

Blood pressure measurement is not suitable in case of serious arteriosclerosis (hardening of the arteries).

- The pulse rate indicator is not suitable for monitoring the frequency of cardiac pacemakers.
- If you suffer from disorders of heart rhythm, known as arrhythmia, you should only use this blood pressure monitor in consultation with your doctor. In certain cases oscillometric measurement methods can produce incorrect readings.
- Pregnant women should only measure their own blood pressure in consultation with their doctor since the readings may be changed by the pregnancy.

Please remember: Self-measurement is not the same as medical treatment! You should never change the dose of medicines as prescribed by your doctor!

Before using your OMRON R7 for the first time

Your blood pressure monitor: the benefits

The OMRON R7 is a compact fully automatic wrist blood pressure monitor. Operating on the oscillometric principle it measures your blood pressure and pulse rate comfortably and quickly from the wrist. It contains an intelligent system for “fuzzy” controlled inflation, known as “Intellisense”. This is an advanced method of oscillometric measurement. No pre-setting of re-inflation is necessary.

The OMRON R7 has a Position Sensor. The Position Sensor is able to determine the optimal height of your wrist (heart level) for the blood pressure measurement. Measuring at the heart level is extremely important since measuring at a higher or lower position will give wrong results.

The OMRON R7 has a large DOT matrix display enabling you to enlarge the digits and showing you graphs. Together with the built in alarm function for same time measurements or pill intake your OMRON R7 is the optimal device to manage your blood pressure.

Main unit




Batteries for monitoring

(Two LR03 alkaline batteries)



Storage case



Instruction Manual



Blood pressure pass



Guarantee card



Names and functions of the parts

Main unit



Systolic Blood Pressure Display

Diastolic Blood Pressure Display

Date/Time Display


Mark

Displayed when a stored measurement is shown.

Pulse Display

◀▶ Button

Push ▶ to move the displayed item forward. Push ◀ to move the displayed item backward.



MEMORY RECALL button

Push this button to see the measured values stored in the monitor.

GRAPH CALL button

Push this button to see the graph.

ON/OFF button

Push this button to measure your blood pressure or to switch the monitor off.


SETTING button

Push this button to set various items on the monitor such as time and date.


How to insert the batteries

Use this monitor after inserting batteries in the main unit.


1. Open the battery compartment.



2. Insert the batteries (Type LR03 / AAA) as shown in the figure. Ensure the polarity (+/-) of the batteries is correct.




3. Close the battery cover.



With new batteries (long life alkaline batteries) you are able to measure app. 300 times under normal conditions. Please switch off the monitor after using it. If you forget to switch the monitor off it will switch off itself after app. 2 minutes. Please take the batteries out of the monitor if you intend not to use the monitor for a long period (app. 2 months).

How to set the Date and Time

Before using your OMRON R7 for the first time, you have to set the date and time. After turning the monitor on by pushing the on/off button the initial screen (see picture) will be available for a few seconds. After the initial screen you will proceed automatically to the TIME SET screen. (For instruction how to set the date and time refer also to page 16 in Chapter How to change the settings.)



Initial Screen



Time Screen

After reaching the time screen you will see the Gregorian calendar (year/day/month). The year will be flashing.

1. By using the (◀ ▶) buttons you can set the current year.
2. By pushing the SET button you will lock the year and proceed to the next item.

For setting the month, date and time repeat step 1 and 2. After setting the minutes you will proceed to the DELETE DATA screen. Please turn off your monitor by pushing the on/off button. Your OMRON R7 is ready for use now.


How to measure Blood Pressure correctly

How to wrap the cuff




In order to measure your blood pressure correctly it is needed to wrap the cuff around your wrist in a correct way. If not, correct measurements are not possible. Please follow the coming instructions (left wrist):

- 1) Make sure that your wrist is bare and will be wrapped over your skin.
- 2) Make sure that your clothing is not constraining the flow of your blood.
- 3) Make sure that the wrist (left or right) to measure is selected from the settings menu (see also page 15). The left wrist is preset.

- 4) Place the cuff over your left wrist with your thumb facing upwards. Leave app 10-15 mm between the cuff and the bottom of the palm of your hand as shown in the figures.



- 5) Pull the wrist cuff upward and close the cuff firmly with the Velcro tape as shown.



You can use your OMRON R7 also on your right hand after selecting right wrist measurement in the SETTING MODE.

People with a slender wrist (wrist circumference of less than 145mm) may feel loose because the cuff may not be wrapped firmly. Please do not force the cuff it will not affect the measurement.



How to obtain a correct posture




Achieving a correct posture before measuring your blood pressure is important since an incorrect posture might give you incorrect readings. To achieve a correct posture please follow the instructions (see pictures):

- 1) Sit down and stretch your back.
- 2) Relax yourself by relieving the stress from your shoulders and arms.
- 3) Breath several times deeply and return to natural respiration.
- 4) Do not sustain your wrist with your other hand and avoid stress in your hand by making firm grips or large extensions.

Correct Posture



Incorrect Posture



Measuring your blood pressure and pulse

If you have achieved the correct posture you can start the measurement. When you measure for the first time the Position Sensor and the Signal will be ON and measuring at the left wrist will be pre set.

Push the on/off button to activate your OMRON R7 and the Position Sensor Screen will be displayed.



How to measure Blood Pressure

Sustain your elbow with your other hand or a table to avoid movement.




• Adjust the height of your wrist by using your elbow as a fulcrum. When the (◀ ▶) will reach the (♥) sign you will hear a beeping sound indicating that your blood pressure monitor is at the correct height (heart height). The (♥) sign will change in the (📶) sign and the measurement will start.



Do not move until the measurement results are displayed!


The screens (see picture) will be shown during the measurement.

When the measurement is finishes, the deflation mark (▼) flashed and the cuff deflates.



How to enlarge the reading

The results will be shown on your display. By pushing the (◀ ▶) buttons you can enlarge the readings.





How to use the Memory function


Your OMRON R7 blood pressure monitor is able to store maximum 90 measurements including date and time. When your blood pressure monitor is holding 90 measurements the oldest will be deleted to store a new measurement.

Press the (M) memory button when your blood pressure monitor is switched off. The latest measurement will be displayed. Each time you push (▶) button an earlier measurement will be displayed. By pushing the (◀) button a later reading is displayed.

By pushing the GRAPH button the selected measurement is displayed in a weekly graph (see also How to use the graph function). Push the memory button to return to the memory display again.



Date and Time of Measurement



How to use the Graph function

This monitor has the possibility to display the stored measurements as a graph. In order to get a clear history of your blood pressure the measurements are put in time zones:

- Morning: Measurements made between 0:00 and 9:59
- Daytime: Measurements made between 10:00 and 17:59
- Evening: Measurements made between 18:00 and 23:59

The last measurement in each time zone will be displayed. For example if you measure 3 times in the same time zone on the same day only the third reading will be displayed. On the other hand all readings will be stored in the memory (See also page 11).


There are three different graphs available:

- All data graph: displays all measurements per week
 - Morning data graph: displays all morning measurements per week
 - Evening data graph: displays all evening measurements per week
- No graph display is provided for daytime data only.

To display the graph push the GRAPH CALL button.




All Data Graph is displayed.



All Data Graph

Push the GRAPH CALL button again to display the Morning Data Graph.

Morning
Data Mark



Morning Data graph

Push the GRAPH CALL button again to display the Evening Data Graph.

Night
Data Mark




Night Data Graph


Each time you push the GRAPH CALL button, All, Morning, and Evening data are displayed in that order repeatedly.

Explanation of the graph display

When the GRAPH button is pushed the latest measurement will appear in a flashing display. In the example the measurement is made on January 31 at 22:40. The graph shows 140 mmHg for systolic and 95 mmHg for diastolic (top and bottom of vertical bar) and a pulse rate of 90 beats a minute.



When you push the memory button in the graph display the details of the measurement will appear. Pushing the graph button again will bring you back the graph display. You can use (◀ ▶) the arrows in the graph and memory display to leaf through the measurements.




Detail Screen of Data in Flashing Display


The systolic blood pressure is higher than 170mmHg

When a systolic blood pressure above 170 mmHg is measured twice the (▲) symbol is displayed in the graph.

When a systolic blood pressure above 170 mmHg is measured three times the (▲) symbol disappears and the screen automatically changes to the "hypertension display". The maximum displayed blood pressure value (y-axis) will change from 170 mmHg to 210 mmHg.



Display screen where the systolic blood pressure values higher than 170 mmHg are measured twice




Graph Display Screen of Hypertension Type

How to change the settings


The settings of your OMRON R7 blood pressure monitor can be changed according to your own preferences and comfort.

To enter the setting menu please press the SET button when your blood pressure monitor is switched off. By using (◀ ▶) you can select your preference. By pressing the SET button again you lock your preference and proceed to the next item as shown in the figure. The on/off button can be used to switch the monitor off again.

- If you push the **SETTING** button while the power is turned off, the Setting Screen is displayed in the order shown in the picture.
- You can select the setting or set the date and time with the buttons.
- To finish the setting, push the on/off button.




Automatic Memory Setting
(Refer to Page 14.)



Position Sensor Setting
(Refer to Page 15.)




Measuring (Wrist) Setting
(Refer to Page 15.)



Alarm Setting
(Refer to Page 15.)




Reset Settings
(Refer to Page 16.)



All Memory Erasure
(Refer to Page 16.)



Time Setting
(Refer to Page 16.)



Auto memory

You can either enable (ON) or disable (OFF) the memory function of your blood pressure. When the auto memory function is switched off the measurement will not be stored. When using your OMRON R7 monitor with more than one person it is recommended to disable the auto memory function. The auto memory function mode is ON when you purchased your OMRON R7 (factory preset).




Automatic Memory
Screen



Automatic Memory
OFF Screen

Position sensor

The position sensor of your OMRON R7 can be switched ON or OFF. In the factory preset the position sensor is enabled. If the position sensor is switched ON you can choose to have the optimal wrist height supported by an acoustic sound (SIGNAL ON or OFF).



Position Sensor Screen




Position Sensor OFF Screen



Position Sensor ON and SIGNAL OFF Screen

Left or right wrist measurement

Normally blood pressure is measured on the left wrist (factory preset). With your OMRON R7 you can choose to use the monitor either on your left or right wrist. It is recommended to use the same wrist when for optimal comparisons of your blood pressure fluctuations.



Measuring (Wrist) Screen



Measuring (Wrist) Right Wrist Screen

Alarm


Your blood pressure monitor has the possibility to set an alarm twice a day. The alarm function can be used to monitor your blood pressure at the same time every day or as a reminder for pill intake.

The alarm will sound for 2 minutes. By pushing the on/off button the alarm will stop. The alarm will not sound when the blood pressure monitor is used (e.g. for measuring or graph display) at the moment the alarm is set.

You can set the requested hour by pushing (◀▶) when flashing. Press SET to lock the hour and proceed with the minutes. By pushing the (◀▶) you can set the minutes. Press SET to confirm and to proceed. Push either (◀▶) to enable (🔔) or disable (🔕) the alarm. Press SET to lock the settings.



Alarm Screen



Screen of indicating that the alarm is sounding

Time


You can change the time and date setting of your blood pressure monitor. When reaching the time screen you will see the Gregorian calendar. The year will be flashing.

1. By using the (◀ ▶) buttons you can set the current year.
2. By pushing the SET button you will lock the year and proceed to the next setting.

For setting the month, date and time repeat step 1 and 2. After setting the minutes you will proceed to the DELETE DATA screen.



Time Screen




All Memory Erasure
YES Screen

Delete data


In this screen you are able to delete all stored measurement from you OMRON R7. Please be careful that deleted data cannot be restored. The factory preset is NO. If you choose YES the monitor will ask for a confirmation "ARE YOU SURE?". To cancel select NO to confirm DELETE DATA select YES.

After deleting all stored measurements the "Data Deleted Screen" will be shown. All the stored measurements are deleted and the monitor will be turned off.

If the stored measurements are not deleted you will proceed to the RESET SETTINGS screen.



Are You Sure YES
Screen




Data Deleted (Blank)
Screen


Reset settings

In this menu you can choose to reset all the settings to the "factory preset" settings. The factory preset for this menu is NO. If you choose YES by using the (◀) button the monitor will ask you for a confirmation "ARE YOU SURE?" After confirmation by selecting YES all the settings will be changed to the factory mode.

The monitor will return to the AUTO MEMORY SCREEN. You can turn off your monitor by pushing the on/off button.



Reset Settings YES
Screen





Are You Sure YES
Screen

What is Blood Pressure?

Blood is pumped to artery by the beat (contraction and expansion) of a pump called the heart. The pressure to pump out blood to artery is called the "systolic blood pressure." The pressure when the heart expands as the blood having circulated the whole body returns is called the "diastolic blood pressure."

Health and blood pressure


Blood pressure fluctuates being affected by everyday life and age. Hypertension is a sign that can cause dangerous diseases such as cerebral apoplexy or myocardial infarction to occur easily. By understanding the state of your blood pressure, review your health and lifestyle.



Hypertension guide

The World Health Organization (WHO) and the International Society of Hypertension (ISH) established the Blood Pressure Classification. Use this Classification as the guide to know the state of your blood pressure. However, if you are concerned with your blood pressure, be sure to follow the instructions of your doctor.

There is no universally accepted definition of hypotension. Generally speaking, those with the systolic pressure below 100 mmHg are considered to be hypotension.



According to the blood pressure classification by the WHO/ISH* (revised in 1999)
* International Society of Hypertension

This classification is made based on the blood pressure values measured at upper arm by sitting on the chair in the hospital regardless of the age.


Blood pressure at the upper arm and wrist

Wrist Blood Pressure Monitor

Blood pressure is usually measured at brachial artery.

As the wrist arterial vessel is connected to the brachial vessel, wrist blood pressure values are close to brachial ones and reflect the change of brachial blood pressures. However in case there is a circulatory disorder such as arterial infarction, fairly large difference is observed between the brachial and the wrist blood pressure values.

After consulting with your doctor, use the wrist blood pressure together with the brachial pressure for your health management.



Diseases influencing blood pressure values

Diabetes, hyperlipemia, or hypertension accelerates arterial sclerosis, which causes dangerous disease such as cerebral apoplexy or myocardial infarction or induces arterial stenosis or peripheral circulatory disorder.

In these cases, the wrist and brachial blood pressure values may differ greatly. (Even in healthy people, a difference of 20 mmHg may occur if the measuring conditions are not proper. Therefore do not try to diagnose the measurement values by yourself, but be sure to follow the instructions of your doctor.)

In addition, difference of wrist and brachial blood pressure values may somewhat get affected by the physiological conditions at the time of measurement. However, the blood pressure values measured at wrist and upper arm usually fluctuate in the same manner. Therefore, you can check the fluctuation tendency of blood pressure by measuring the wrist blood pressure.



Differences between blood pressure measured at home and at the doctor's office

Normally blood pressure measured at home tends to be lower than when measured at the doctor's office. This can be as much as 20 mmHg due to that you feel more relaxed at home and feel slightly tense when visiting the doctor. To manage your blood pressure it is important to monitor and register your blood pressure at home.

Since blood pressure is fluctuating from beat to beat and is influenced by daily activities measurements can differ from measurement to measurement.

Maintenance and storage






- For cleaning the monitor only use a soft, lightly moistened cloth.
- **Do not use petroleum spirits, thinners or similar solvents!**
- Stains on the cuff can be carefully removed using a moist cloth and soapy water.
- **Do not wash the cuff or make it wet. Please make sure no water will enter the main unit!**
- Keep your monitor in the storage case protected from extreme temperatures and humidity.
- Do not carry out repairs of any kind yourself. If a defect occurs consult your OMRON dealer or distributor.
- Please take the batteries out if the monitor is not used for a long period (app. 2 months or more).
- Electro Magnetic fields can influence the measurement. Do not use a portable phone near the unit.

Disposal of this product and used batteries should be carried out in accordance with the national regulations for the disposal of electronic products.



Error indication

If the measurement is not made correctly, the following error can be displayed.

Error	Cause	How to correct
 <p>After several seconds</p>	<p>You moved your wrist during measurement.</p>	<p>Measure your blood pressure again without moving your wrist until the measurement is completed. (Refer to Page 10.)</p>
 <p>After several seconds</p>	<p>You moved your arm or body, or talked during measurement.</p>	<p>Measure your blood pressure again without moving your arm or body. (Refer to Page 10.)</p>
 <p>After several seconds</p>	<p>You moved your arm or body extensively during measurement.</p>	<p>Measure your blood pressure again without moving your arm or body. (Refer to Page 10.)</p>
 <p>Battery replacement mark flashes.</p>	<p>Batteries are worn out.</p>	<p>Replace both batteries with new ones. (Refer to Page 6.)</p>
 <p>Only E and numerals are displayed.</p>	<p>Trouble caused by abnormality in function.</p>	<p>Please contact the nearest OMRON dealer.</p>

If you cannot resolve the problem even after taking the above-mentioned measures, please contact the nearest OMRON dealer.

Troubleshooting

Trouble	Possible cause	How to correct
Nothing is displayed when you push the on/off button.	Batteries are worn out.	Replace the batteries with new ones. (See to Page 6.)
	The directions of polarities of inserted batteries (+ and -) are wrong.	Insert the batteries correctly. (Refer to Page 6.)
Measurement cannot be made. Or the blood pressure values are displayed abnormally high (low).	The wrist cuff is not held at the level of the heart.	Measure your blood pressure again in correct posture. (Refer to Page 9.)
	The wrist cuff is not wrapped firmly.	Wrap the wrist cuff correctly. (Refer to Page 7.)
	Too much stress is applied on your shoulder or arm.	Relax yourself and make measurement.
	You are talking or moving your hands during measurement.	Be still and make measurement.
The blood pressure values differ from the values measured at brachial.	Refer to the section on "Wrist Blood Pressure Monitor." (Refer to Page 18.)	
Blood pressure values are displayed before the monitor inflates to the required value.	Air is leaking from the wrist cuff.	Contact the store where the unit was sold or the nearest OMRON distributor.
Blood pressure value fluctuates each time of measurement. Or the blood pressure values are displayed abnormally high (low).	Blood pressure value tends to fluctuate depending on the mental state during measurement or the measuring time. Breathe deeply to relax yourself and make measurement again. (Refer to Page 19.)	

If you cannot resolve the problem even after taking the above-mentioned measures, please contact the nearest OMRON dealer.

Spare parts and accessories

OMRON considers, due to the used high company and quality standards, the main unit as a non-serviceable part because of the necessity of proper calibration after replacement of high-tech components.

The cuff can be replaced by an authorised expert.

Specifications

Product Description :	Automatic Blood Pressure Monitor
Model :	R7 (HEM-637-E2 / HEM-637-E7)
Display :	Dot matrix LCDdisplay
Measurement :	Oscillometric method
Measurement Range :	Pressure; 0 to 299 mmHg, Pulse rate; 40 to 180 /min.
Accuracy :	Pressure; within +/-3 mmHg, Pulse rate; within +/-5% of
reading	
Inflation :	Automatic inflation with pump
Deflation :	Automatic rapid deflation
Power supply :	Two LR03 alkaline batteries (AAA)
Battery Life :	Approximately 300 times of measurement when using alkaline batteries (In case of the monitor is used three times a day inflating to 170 mmHg at the room temperature of 22°C)
Operating Temperature and Humidity :	+10° to +40°C, 30 to 85% RH
Storage temperature and humidity :	-20° to +60°C, 10 to 95% RH
Measurable circumference of wrist :	135 to 215 mm
Weight of Main Unit :	Approximately 150 g (not including batteries)
External Dimensions :	78 (W) x 65.5 (H) x 37.2 (D) mm (not including the wrist cuff)
Electric Shock Protection :	Internal power supply appliance type B
Accessories :	Two LR03 alkaline batteries for monitoring use, storage case. Instruction Manual, blood pressure pass, guarantee card.

• In the interest of product improvement, specifications are subject to change without notice.

Manufacturer	OMRON HEALTHCARE Co., Ltd. 24, Yamanouchi Yamanoshita-cho, Ukyo-ku, Kyoto, 615-0084 JAPAN
EU-Representative:	OMRON HEALTHCARE EUROPE B.V. Kruisweg 577, 2132 NA Hoofddorp, THE NETHERLANDS Made in China

This device fulfils the provisions of the EC directive 93/42/EEC (Medical Device Directive). This blood pressure monitor is designed according to the European standard EN1060, Non-invasive sphygmomanometers Part 1: General Requirements and Part 3: Supplementary requirements for electromechanical blood pressure measuring systems.

This OMRON product is produced under the strict quality system of OMRON HEALTHCARE Co., Ltd., Japan. The Core component for OMRON blood pressure monitors, which is the Pressure Sensor, is produced in Japan.



=Type B

CE 0197

Manufacturer 	OMRON HEALTHCARE Co., Ltd. 53, Kunotsubo, Terado-cho, Muko, Kyoto, 617-0002 JAPAN			
EU-representative <table border="1" data-bbox="76 268 252 339"> <tr> <td data-bbox="76 268 163 339">EC</td> <td data-bbox="163 268 252 339">REP</td> </tr> </table>	EC	REP	OMRON HEALTHCARE EUROPE B.V. Scorpius 33, 2132 LR Hoofddorp THE NETHERLANDS www.omron-healthcare.com	
EC	REP			
Production facility Produktionsstätte Productiefaciliteit Site de production Oficina de producción Stabilimento di produzione	OMRON (DALIAN) CO., LTD. Dalian, CHINA			
Subsidiary Niederlassung Dochteronderneming Succursale Empresa filial Consociata	<table border="1" data-bbox="439 589 970 1138"> <tr> <td data-bbox="439 589 970 701"> OMRON HEALTHCARE UK LTD. Opal Drive, Fox Milne, Milton Keynes MK15 0DG, U.K. </td> </tr> <tr> <td data-bbox="439 701 970 875"> OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim GERMANY www.omron-medizintechnik.de </td> </tr> <tr> <td data-bbox="439 875 970 1138"> OMRON SANTÉ FRANCE SAS 14, rue de Lisbonne 93561 Rosny-sous-Bois Cedex, FRANCE Uniquement pour le marché français: OMRON Service Après Vente N° Vert 0 800 91 43 14 consommateurs@omron-sante.fr www.omron-sante.fr/contact </td> </tr> </table>	OMRON HEALTHCARE UK LTD. Opal Drive, Fox Milne, Milton Keynes MK15 0DG, U.K.	OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim GERMANY www.omron-medizintechnik.de	OMRON SANTÉ FRANCE SAS 14, rue de Lisbonne 93561 Rosny-sous-Bois Cedex, FRANCE Uniquement pour le marché français: OMRON Service Après Vente N° Vert 0 800 91 43 14 consommateurs@omron-sante.fr www.omron-sante.fr/contact
OMRON HEALTHCARE UK LTD. Opal Drive, Fox Milne, Milton Keynes MK15 0DG, U.K.				
OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH John-Deere-Str. 81a, 68163 Mannheim GERMANY www.omron-medizintechnik.de				
OMRON SANTÉ FRANCE SAS 14, rue de Lisbonne 93561 Rosny-sous-Bois Cedex, FRANCE Uniquement pour le marché français: OMRON Service Après Vente N° Vert 0 800 91 43 14 consommateurs@omron-sante.fr www.omron-sante.fr/contact				

Made in China
Hergestellt in China
Geproduceerd in China
Fabriqué en Chine
Fabricado en China
Prodotto in Cina